



STERILIZATION LOG SHEET

Month/Year: _____

Location/Unit: _____

Date (dd/mm/yy)	Time			Temp. °F or °C	Pressure	Temperature-sensitive indicator: Colour Change Observed (Y/N)	Operator's Initials	Comments
	Start	End	Cycle Length					

Biweekly spore strip tests submitted: (dd/mm/yy) Result received (dd/mm/yy) Result received (dd/mm/yy) Result received

PLEASE NOTE:
 1. Pressure cookers, glass-bead sterilizers, microwaves, U.V. sterilizers, boiling water and dry-heat domestic ovens are **NOT** approved means of sterilizing.
 2. The manufacturer's instructions for the sterilizing equipment must be onsite and easily accessible.
 3. A record book for monitoring each load placed in the sterilizer (temperature, pressure and cycle length) must be maintained for a minimum of five years, with the last 12 months onsite.
 4. Biweekly spore strip test results must be kept for a minimum of five years, with the last 12 months onsite.
 5. Failed spore strip tests – refer to *Biological Monitoring of Sterilizer (Spore Test)* poster.

Sources: Ministry of Health and Long Term Care, Infection Prevention and Control Best Practices for Personal Services Settings, Jan. 2009.